

# Reference Material Data Sheet

## *1C3 Cigarillo*

**Data Sheet Number: 2022-1C3CTRP**

Reference values generated on: 08/01/2022

Reference values are valid until: 08/01/2027

Superseded data sheet code: NA

Description of Material	Machine made cigarillo
Lot/Batch Number	<b>1C3 Small Machine Made Cigarillo</b>
Matrix	Tobacco blend

### Notes

- Reference values listed below reflect analysis results submitted during CORESTA studies (References below) satisfying Clause 5.15 of ISO Guide 34:2009 and statistically evaluated using Approach A in Clause 10.5 of ISO Guide 35:2017.
- The "Reference Uncertainty" listed herein are expanded uncertainties obtained by multiplying the combined standard uncertainty by a constant coverage factor of 3, i.e.  $k = 3$ .
- The reference values listed herein are reflective of chemical analysis done on a "as is" basis, i.e. the samples were not dried prior to analysis.
- Cigars were smoked using the CORESTA Recommended Method No. 64.
- The values will be updated by the valid until date or as needed.
- Due to the possibility of self-extinguishing, relighting the cigar may be necessary.

### *1C3 - Small Machine Made Cigarillo Values and Uncertainties*

Mainstream Smoke					
Parameter	Reference Value	Certified Uncertainty	Unit	Number of Accepted Data Points	Constant Coverage Factor
Total Particulate Matter (TPM)	59.44	14.87	mg/cigar	45	3
Tar (Nicotine-free dry particulate Matter (NFDPM))	53.53	14.07	mg/cigar	45	3
Carbon Monoxide (CO)	94.06	36.69	mg/cigar	45	3
Nicotine	1.63	1.14	mg/cigar	45	3
Puff Count	35.94	5.28	puffs/cigar	45	3
Water	4.24	2.06	mg/cigar	45	3
Weight	2747.4	91.5	mg/cigar	45	3

Homogenized Cigar					
Parameter	Reference Value	Certified Uncertainty	Unit	Number of Accepted Data Points	Constant Coverage Factor
Ammonia	2869	1028	µg/g	36	3
Arsenic	587	762	ng/g	24	3
Cadmium	1101	639	ng/g	24	3
Nicotine	6.97	0.89	mg/g	33	3
NNK	694	152	ng/g	36	3
NNN	4264	1418	ng/g	36	3
NAT	1659	408	ng/g	36	3
NAB	158	46	ng/g	36	3
Moisture Content	15.48	3.05	%	33	3
pH	6.39	0.28	pH units	39	3
Water Activity	0.66	0.11	a <sub>w</sub>	30	3

---

**CERTIFIED VALUES AND UNCERTAINTIES:**

The Certified Values and Uncertainties for the mainstream smoke parameters contained herein are reflective only of data obtained from a Linear Smoking machine.

**DATA AQUISITION:**

Data used in calculations for constituents were reported in multiple studies listed below.

Mainstream smoke – CORESTA Technical Report Cigar Smoke Analysis 14<sup>th</sup> Collaborative Study 2019-2020 (April 2021)

Homogenized cigar – CORESTA Technical Report University of Kentucky Cigar Reference Products 2021 Analysis (September 2021)

**INTENDED USE:** This product may be used for analytical method development, assigning values to materials (when applicable), and, equipment calibration to any applicable extent.

**INSTRUCTIONS FOR CORRECT USE:** This reference material may be stored in sealed containers or bags at -20°C until testing. Prior to analysis, the reference material should be unopened and transferred to a refrigerator for a minimum of 24 hours or until they are completely thawed, and then moved to room temperature for at least 2 hours until it reached temperature equilibrium. Once the reference materials are equilibrated to room temperature, the reference materials may be stored at 4°C for up to one week, if they will not be analyzed immediately. Any reference materials found to have damaged package should be discarded.

**HAZARD INFORMATION:** N/A

**HOMOGENEITY:** Homogeneity of this material is reflected in the “Certified Uncertainties” disclosed herein.

---

**APPROVING PERSONNEL:** This material is approved by **Huihua Ji, C. Ruth McNees** and **Ling Yuan** on behalf of CTRP