

**CIG-2020B Proficiency
Program Protocol
2020 Round 2**

Objective

This round of testing will include smoking the 1R6F reference cigarette using both the International Organization for Standardization (ISO) and the Health Canadian Intense (HCI) regimes. **Please do not round results at any point in your calculations.**

The mainstream smoke measured properties are:

- Nicotine-Free Dry Particulate Matter (NFDPM/Tar)
- Nicotine
- Carbon Monoxide
- Water
- Total Particulate Matter (TPM)
- Puff Count
- Benzo [a] pyrene (BaP)
- BaP-Total Particulate Matter (BaP-TPM)
- BaP-Puff Count

Proficiency Study Timeframe

Due to the COVID-19 pandemic, dates are subject to change. Please check ctrp.uky.edu for updates during the study. To request an extension, please contact ruth.mcnees@uky.edu or ctrp@uky.edu.

April 16, 2020: 09:00 AM EDT

PT round Opens, Test Kits available for purchase

April 30, 2020: 09:00 AM EDT

Data submission portal Opens, First day of data submission

July 9, 2020: 5:00 PM EDT

Data submission portal Closes, Final day of data submission

July 30, 2020

Target date for issuance of Interim Report

August 27, 2020

Target date for issuance of Final Report, PT round Closes

Eastern Daylight Time (EDT) (New York, NY time)

Test kits are available for purchase beginning on **April 16, 2020**. This round of testing for data submission will open on **April 30, 2020** and close on **July 9, 2020**. The University of Kentucky, Center for Tobacco Reference Products (CTRP) data submission portal will be locked after the closing date and will no longer accept results. Results obtained after the closing date will not be included in the proficiency study report. The target date for issuance of the interim report is **July 30, 2020**. The participants are encouraged to review the interim report and provide feedback, i.e. comments, erroneous data entry, additional notes, etc., through the online feedback form

available by clicking “Submit Comments” next to the interim report link located on the “My Proficiency Studies” tab of the CTRP website (ctrp.uky.edu). Feedback received will be considered and, if appropriate, incorporated in a final report which will be issued, tentatively, on **August 27, 2020**. The interim report and final report can be downloaded from the “My Proficiency Studies” tab located on the CTRP website by clicking the “Interim Report” or “Analysis Report” link, respectively.

References

Conditioning	ISO 3402:1999
TPM/ NFDPM	ISO 4387:2019
Carbon Monoxide	ISO 8454:2007
Nicotine	ISO 10315:2013
Water (GC)	ISO 10362-1:2019
Benzo [a] pyrene	ISO 22634:2017, CORESTA N 58
Intense smoking regime, puff parameters	Health Canada Method T-115
ISO smoking regime, puff parameters	ISO 3308:2012, ISO 4387:2019

Proficiency Test Material (1R6F Reference Cigarettes)

Proficiency Test Material for this round of proficiency sampling must be obtained from the CTRP by procuring the proficiency test kit. The materials will come with a test protocol and instructions to download the electronic reporting template in the form of a pre-formatted MS Excel file. It is not acceptable to use 1R6F reference cigarettes from your inventory. Using the materials provided will ensure that all participants are using cigarettes from a batch that is pre-characterized for the purposes of the Proficiency Test Scheme. Homogeneity of the Proficiency Test Material was determined by selecting 12 random samples and having them analyzed in at least triplicate. The testing was sub-contracted to a third-party laboratory meeting the quality requirements of the proficiency testing scheme in accordance with ISO/IEC 17043. Test results confirm that the Proficiency Test Material is fit for proficiency testing.

General Guidance

Table 1 lists the smoking parameter and vent blocking specifications for each smoking regimen. The butt length for this testing is set at 35 mm or 1.38 inches.

It is important to note the need for participants to record any deviation from the standard methods in their report. Operating conditions considered optional reporting by the laboratory should also be recorded on their report. Any circumstances that arise during the analysis of these cigarettes which may influence either the precision or the bias of the result should be recorded in the report. Details of deviation from normal operations should be recorded in the “Notes” section of the Excel reporting template.

Based on historical data from previous rounds of testing, we anticipate participation from approximately 20 laboratories for this Proficiency Testing Scheme. Any participant that does not receive a proficiency testing kit or receives a damaged kit is encouraged to contact the CTRP (ctrp@uky.edu) immediately to ensure that the participant has sufficient time to complete the Proficiency Test Scheme as scheduled.

Participants should confirm the type of smoking machine being used (rotary or linear), and report the model and manufacturer. The temperature and relative humidity at the time the smoking is conducted should be recorded.

Table 1
Smoking parameter specifications

<i>Smoking Regimen</i>	<i>Puff Volume (mL)</i>	<i>Puff Interval (s)</i>	<i>Puff Duration (s)</i>	<i>Vent Blocking</i>
ISO	35± 0.3	60± 0.5	2 ± 0.02	0%
HCI	55 ± 0.5	30 ± 1	2 ± 0.02	100%

Note: Puff Interval is time in seconds from the start of one puff to the start of the next puff.¹

Test Item Storage

The samples should be stored in plastic bags at 4°C prior to conditioning for the proficiency test.

Conditioning

Samples should be conditioned for a minimum of 48 hours, but no more than 10 days at 22 ± 1 °C and 60 ± 3% relative humidity before conducting each smoking and physical parameter test.

Replicates Required

Replicates must be obtained **under repeatability conditions** i.e. same instrument and same operator.

Smoking Parameter

Smoke 5 replicates for each smoking regime for both linear and rotary smoking machines. Laboratories should follow their routine smoking plan. Expected values, and the metrological traceability and uncertainty, for selected analytes can be found in the Certificate of Analysis for the 1R6F certified reference cigarette at the CTRP website (ctrp.uky.edu). Results will be presented as consensus values based on participant results.

Please report on as many analytes as you can.

Participants who order a Proficiency Test Kit should download the Excel reporting template which will be used to submit results for the proficiency testing. The Excel reporting template can be downloaded from the “My Proficiency Studies” tab located on the CTRP website (ctrp.uky.edu) after you have purchased a Proficiency Test Kit. If you order a linear and a rotary Proficiency Test Kit or multiple kits for each machine, you must download the Excel reporting template for each of the Proficiency Test Kits. Each Excel reporting template has a unique “Assigned Data Set ID” based on a customer’s purchase. Please make sure that you enter the data into the correct Excel reporting template (linear or rotary). Please note that there is no need to round results at any point in your calculations. Make sure to report results in the units indicated in the Excel reporting template. The results should be submitted electronically through the CTRP

¹ 2012 CORESTA Collaborative Study for CORESTA Monitor #7 (CM7) for Determination of Test Piece, Weight, TPM, Water, Nicotine, NFDPM, Carbon Monoxide and Puff Count Obtained Under Mainstream ISO and HCI Smoking Regimes; May 2013

website on the “My Proficiency Studies” tab. The participant will: (1) click the blue “Submit Proficiency Data” button for the correct reporting proficiency study; (2) browse their computer for the Excel reporting template for that proficiency study; (3) select the appropriate file; and (4) then click the “Load and Review Data” button. The participant will have the opportunity to review their data online before their final submission of data to the CTRP. Participants are encouraged to provide the data collected for each round of testing without discussing the results with other potential participants.

The study report will contain:

- Executive Summary
- Purpose of study
- Protocol
- Coded laboratory raw data
- Statistical summary (both graphical and numeric)

File Formatting Requirements for Data

To ensure clear and uninterrupted data processing among disparate computer systems, please use the Excel reporting template provided with the Proficiency Test Kit, which has been formatted for data entry. Please note that the downloadable Excel reporting template contains “locked” codes and a Proficiency Study ID (CIG-2020B) and an Assigned Data Set ID number specific to your test kit and this round of Proficiency Testing.

Common sources of data error include, but are not limited to, incorrect units for reporting data (mg/cig instead of g/cig), failure to calculate values for individual cigarettes, or improper calibration.

Below is a description of the file formatting, type, and expected contents of data files to be sent to the CTRP Proficiency Testing Program.

File Details

The data transport file should be formatted as an Excel file, specifically the XML-based format that Excel files are saved in by default. There should be no spaces in the filename. The Excel file extension should, by default, be .xlsx. **Example:** *linear_datasetid_3476.xlsx*

Proficiency Data

Please use the dropdown box in the top right section of the Excel reporting template to answer whether the lab has ISO Accreditation.

Machine Smoking Data

Please be sure to enter data for the specific smoking machine (linear or rotary) used in your analysis:

- *Smoking Machine Make (i.e., manufacturer)*

- *Smoking Machine Model*
- *Enter any notes on data collection (if necessary)*

ISO Data - TNCO

Please be sure to enter measurements for the specific ISO Data:

- *ISO Data Test Date*
- *Linear Machines: Enter the number of ports used per replicate.
Rotary Machines: Enter the number of collections used per replicate.*
- *Linear Machines: Enter the number of cigarettes smoked per port.
Rotary Machines: Enter the number of cigarettes per collection.*
- *Laboratory conditions (6 variables) for each of replicates*
- *Measurements for each of the 4 smoking parameters (Nicotine-Free Dry Particulate Matter, Nicotine, Carbon Monoxide, and Water) for each of the 5 replicates.*
- *Please use the dropdown menu to select the “instrument” used in testing for each of the 5 smoking parameters.*
- *Please use the dropdown menu to select the “method” used in testing for each of the 5 smoking parameters. If your method is not identified in the dropdown menu, please type your method in the box provided.*

	Nicotine	Water
Instrument	GC/MS or GC/FID	Karl Fischer Apparatus or GC
Method 1	CRM#, ISO #, or in-house	CRM#, ISO #, or in-house
Enter alternate method (text)		

For the instrument: GC/MS- Gas Chromatography/mass spectrometers, FID- flame ionization detector

For method: CRM: CORESTA Recommended Method; in-house method or others

- *Please type your internal standard for the **nicotine** and **water** smoking parameters in the box provided.*

	Nicotine	Water
Internal Standard (enter text)		

- *Measurements for Total Particulate Matter and Puff Count.*
- *If a participant does not have a measurement for a data field, please leave the Excel cell blank. When you upload your spreadsheet to the CTRP database, you will have the opportunity to review all your data on the web-based user interface. All cells that were left blank on the Excel reporting template, will appear as “< empty >” on the web-based user interface screen.*

ISO Data – Benzo[a]pyrene (BaP)

Please be sure to enter measurements for the specific ISO Data:

- *ISO Data Test Date*
- *Linear Machines: Enter the number of ports used per replicate.*
Rotary Machines: Enter the number of collections used per replicate.
- *Linear Machines: Enter the number of cigarettes smoked per port.*
Rotary Machines: Enter the number of cigarettes per collection.
- *Laboratory conditions (6 variables) for each of replicates*
- *Measurements for BaP of the 5 replicates.*
- *Please use the dropdown menu to select the “instrument” used in testing for BaP measurement.*
- *Please use the dropdown menu to select the “method” used in testing for BaP smoking parameter. If your method is not identified in the dropdown menu, please type your method in the box provided.*

	BaP
Method 1	GC/MS
Method 2	LC/FLD
Enter alternate method (text)	

LC/FLD: Liquid Chromatography-Fluorescence Detector

GC/MS: Gas chromatography–mass spectrometry

- *Please use the dropdown menu to select the “internal standard” used in testing for BaP smoking parameters. If your method is not identified in the dropdown menu, please type your method in the box provided.*

	BaP
Internal Standard Option 1	BaP-d12
Other (enter text)	

- *Measurements for BaP Total Particulate Matter and BaP Puff Count.*
- *If a participant does not have a measurement for a data field, please leave the Excel cell blank. When you upload your spreadsheet to the CTRP database, you will have the opportunity to review all your data on the web-based user interface. All cells that were left blank on the Excel reporting template, will appear as “< empty >” on the web-based user interface screen.*

Intense Smoking Data - TNCO

Please be sure to enter measurements for the specific Intense Smoking Data:

- *Intense Test Date*
- *Linear Machines: Enter the number of ports used per replicate.*
Rotary Machines: Enter the number of collections used per replicate.

- *Linear Machines: Enter the number of cigarettes smoked per port.
Rotary Machines: Enter the number of cigarettes per collection.*
- *When smoking is done using the HCl smoking regime, please be sure to enter data for the Ventilation Blocking Method*
- *Laboratory conditions (6 variables) for each of replicates*
- *Measurements for each of the 4 smoking parameters (Nicotine-Free Dry Particulate Matter, Nicotine, Carbon Monoxide, and Water) for each of the 5 replicates.*
- *Measurements for Total Particulate Matter and Puff Count.*
- *If a participant does not have a measurement for a data field, please leave the Excel cell blank. When you upload your spreadsheet to the CTRP database, you will have the opportunity to review all your data on the web-based user interface. All cells that were left blank on the Excel reporting template, will appear as "< empty >" on the web-based user interface screen.*

Intense Smoking Data – Benzo[a]pyrene (BaP)

- *Intense Test Date*
- *Linear Machines: Enter the number of ports used per replicate.
Rotary Machines: Enter the number of collections used per replicate.*
- *Linear Machines: Enter the number of cigarettes smoked per port.
Rotary Machines: Enter the number of cigarettes per collection.*
- *When smoking is done using the HCl smoking regime, please be sure to enter data for the Ventilation Blocking Method*
- *Laboratory conditions (6 variables) for each of replicates*
- *Measurements for BaP of the 5 replicates.*
- *Measurements for BaP Total Particulate Matter and BaP Puff Count.*
- *If a participant does not have a measurement for a data field, please leave the Excel cell blank. When you upload your spreadsheet to the CTRP database, you will have the opportunity to review all your data on the web-based user interface. All cells that were left blank on the Excel reporting template, will appear as "< empty >" on the web-based user interface screen.*

File Data

Additional information/instructions are available on the home page of the CTRP website in the document section at [How to Upload Proficiency Test Data](#) .

Completed Files

When the Excel reporting template is completed and saved with the current date embedded in the file name, please submit the data file through the "My Proficiency Studies" tab located on the CTRP website following the instructions set forth above. The data will be stored anonymously, based on a randomly generated Assigned Data Set ID in a secured database for the study. All data will be treated in a confidential manner as set forth in the "Terms and Conditions for CTRP Proficiency Testing Programs," and agreed to by the participants.

Statistical Analysis

Estimates of the robust mean and robust standard deviation from applying Algorithm A within and between labs will be used for the computation of the repeatability standard deviation (s_r) and reproducibility standard deviation (s_R).

Statistical evaluation of a participants' performance will be based on the number of datasets submitted for each analyte. If a sufficient number of datasets are submitted, the lab's performance will be scored based on Z-scores, $Z = \frac{x_i - x_{pt}}{\sigma_{pt}}$, where x_i is the robust mean of participant for a given measurand, x_{pt} is the assigned value for the proficiency test, and σ_{pt} . The Z-scores are commonly interpreted as

- | | | |
|-------|-------------------|--|
| (i) | $ Z \leq 2.0$ | Satisfactory, acceptable |
| (ii) | $2.0 < Z < 3.0$ | Questionable, a warning signal (W) is given |
| (iii) | $ Z \geq 3.0$ | Unsatisfactory, an action signal (A) is given. |

If an insufficient number of datasets are submitted for rigorous statistical evaluation, alternative scoring methods discussed in ISO 13528 will be employed and details will be provided in the interim report for participants to review and comment. The final report will contain the detailed approach for scoring performance.

Proficiency Test Contacts

The Logistics Coordinator and Quality Manager for the Proficiency Testing Program for the Center for Tobacco Reference Products (CTRP) are listed in this section.

CTRP - Kentucky Tobacco & Research Development Center
1401 University Drive
Lexington, KY 40546-0236
CTRP@uky.edu

For logistics (shipping, customs, etc.) concerns please contact the Logistics Coordinator:

James T. Hall
Kentucky Tobacco & Research Development Center
1401 University Drive, Room B07
Lexington, KY 40546-0236
Phone: 859-257-2660
james.hall@uky.edu

For analytical or reporting concerns please contact the Quality Manager:

Ruth McNees
Kentucky Tobacco & Research Development Center
1401 University Drive, Room 200E
Lexington, KY 40546-0236
Phone: 859-257-9133
Ruth.mcnees@uky.edu