

University of Kentucky Center for Tobacco Reference Products

1401 University Drive Lexington, KY 40546-0236 www.ctrp.uky.edu Telephone: (859) 257-2660 Fax: (859) 257-6012 Email: <u>ctrp@uky.edu</u>

Reference Material Data Sheet

High TSNA Ground Tobacco

Data Sheet Number: 2016-GT01CTRP

Reference values generated on: 08/01/2016 Reference values are valid until: 08/01/2017 Superseded data sheet code: N/A

| Description of Material | Ground raw tobacco |
|------------------------------------|------------------------|
| Lot/Batch Number | RT5 |
| Matrix Major starting materials | Single-variety tobacco |
| Major starting materials | Dark air-cured tobacco |

Notes

- 1. Reference values listed below reflect analysis results submitted by a single laboratory using a single method satisfying Clause 5.15 of ISO Guide 34:2009 and statistically evaluated using Approach A in Clause 10.5 of ISO Guide 35:2006.
- 2. The "Reference Uncertainty" listed herein are expanded uncertainties obtained by multiplying the combined standard uncertainty by a constant coverage factor of 3, i.e. *k* = 3.
- 3. The reference values listed herein are reflective of chemical analysis done on a "as is" basis, i.e. the samples were not dried prior to analysis.

RT5 Reference Values and Uncertainties

| Reference Values and Uncertainties | | | | | | | |
|------------------------------------|--------------------|--------------------------|------|--------------------------------------|--------------------------------|--|--|
| Parameter | Reference Value | Reference Uncertainty | Unit | Number of accepted data points | Constant Coverage factor | | |
| Nicotine | 28280 | 236 | μg/g | 20 | 3.00 | | |
| Nornicotine | 730 | 19 | μg/g | 20 | 3.00 | | |
| Anabasine | 130 | 6 | μg/g | 20 | 3.00 | | |
| Anatabine | 648 | 18 | μg/g | 20 | 3.00 | | |
| NNN | 7.01 | 0.14 | μg/g | 20 | 3.00 | | |
| NAT | 4.04 | 0.10 | μg/g | 20 | 3.00 | | |
| NAB | 0.43 | 0.03 | μg/g | 20 | 3.00 | | |
| NNK | 0.93 | 0.05 | μg/g | 20 | 3.00 | | |
| Moisture | 12.6 | 0.2 | % | 20 | 3.00 | | |

INSTRUCTIONS FOR CORRECT USE:

The reference material should be stored in the original packaging or in airtight containers just large enough to contain the sample in a cool dry place ($^{4}\circ$ C) upon receipt. If the reference material will not be used within 1 week , it is recommended that the reference material be stored at, or below, -20° C until needed. Allow the reference materials to thaw and equilibrate in the refrigerator ($^{4}\circ$ C) for 24 hours and then at least 1 hour at ambient conditions prior to using.

HAZARD INFORMATION: N/A

HOMOGENEITY: Homogeneity of this material is reflected in the expanded uncertainties disclosed herein.

APPROVING PERSONNEL: This material is approved for distribution by **Huihua Ji** and **Socrates Jose P. Cañete** on behalf of CTRP.